

MaxQ 2000 CO₂ Plus

Digital CO₂ Resistant Shaker

Catalog #88881101, #88881102

#88881103, #88881104

Operation Manual 0000847

Visit us online to register your warranty
www.thermoscientific.com/labwarranty

Thermo
SCIENTIFIC

MANUAL NUMBER 0000-847

1	--	10/28/21	88881127B and 88881181 are introduced.	TFI
0	0	18/11/2020	Original	
REV	ECR/ECN	DATE	DESCRIPTION	By

Do You Need Information or Assistance on Thermo Scientific Products?

If you do, please contact us 8:00 a.m. to 6:00 p.m. (Eastern Time) at:

1-740-373-4763

1-800-438-4851

1-877-213-8051

<http://www.thermofisher.com>

service.led.marietta@thermofisher.com

www.unitylabservices.com

Direct

Toll Free, U.S. and Canada

FAX

Internet Worldwide Web Home Page

Tech Support Email Address

Certified Service Web Page

Our **Sales Support** staff can provide information on pricing and give you quotations. We can take your order and provide delivery information on major equipment items or make arrangements to have your local sales representative contact you. Our products are listed on the Internet and we can be contacted through our Internet home page.

Our **Service Support** staff can supply technical information about proper setup, operation or troubleshooting of your equipment. We can fill your needs for spare or replacement parts or provide you with on-site service. We can also provide you with a quotation on our Extended Warranty for your Thermo Scientific products.

Whatever Thermo Scientific products you need or use, we will be happy to discuss your applications. If you are experiencing technical problems, working together, we will help you locate the problem and, chances are, correct it yourself...over the telephone without a service call.

When more extensive service is necessary, we will assist you with direct factory trained technicians or a qualified service organization for on-the-spot repair. If your service need is covered by the warranty, we will arrange for the unit to be repaired at our expense and to your satisfaction.

Regardless of your needs, our professional telephone technicians are available to assist you Monday through Friday from 8:00 a.m. to 6:00 p.m. Eastern Time. Please contact us by telephone or fax. If you wish to write, our mailing address is:

Thermo Fisher Scientific (Asheville) LLC
401 Millcreek Road, Box 649
Marietta, OH 45750

International customers, please contact your local Thermo Scientific distributor.

Table of Contents

Section 1	Unpacking and Installation	1-1
	Packing List	1-1
	Universal (Aluminum) Platform Installation	1-3
	18 inch (Aluminum) Platform Installation	1-4
	Rubber Mat Platform Installation	1-5
	Connections of the Instrument	1-6
Section 2	Overview	2-1
	Specifications	2-1
	Environmental Conditions	2-2
	Safety Tips	2-2
	Storing the Signal Cable	2-3
	Output Current of Control Box	2-3
	Maximum Shaking Speed	2-5
Section 3	Operation	3-1
	Operations on the Control Panel	3-1
	Speed and Time Settings	3-2
	Alarm System	3-3
Section 4	Application	4-1
Section 5	Safety Tips	5-1
Section 6	Optional Accessories	6-1
Section 7	Troubleshooting	7-1
Section 8	Warranty Information	8-1

Section 1 Unpacking and Installation

1. Inspect contents upon receipt of the instrument, if the instrument is upside down, contact the manufacturer immediately.
2. Check packing list below. If there are any damaged or missing items, contact the manufacturer.
3. Place the instrument on a level and firm surface, to avoid vibration and noise.

Table 1-1. Packing List













Thermo Scientific Cat. No.	88881101	88881102	88881103	88881104	Pictutre
Model name on carton box label.	MaxQ 2000 CO2 Plus, 100-240V, for USA	MaxQ 2000 CO2 Plus, 100-240V, various plugs	MaxQ 2000 CO2 Plus Promotion Package for USA, includes CO2 resistant shaker (88881101), Universal Platform and Flask Clamp Starter Kit	MaxQ 2000 CO2 Plus Promotion Package for EMEA & APAC, includes CO2 resistant shaker (88881102), Universal Platform & Flask Clamp Starter Kit	
Main Body	1	1	1	1	
Control Box	1	1	1	1	
General Power Adaptor	1	1	1	1	
US Plug	1	N/A	1	N/A	
CN Plug	N/A	1	N/A	1	

Table 1-1. Packing List (Continued)

Thermo Scientific Cat. No.	88881101	88881102	88881103	88881104	Pictutre
Model name on carton box label.	MaxQ2000C02 Plus, 100-240V, for USA	MaxQ2000C02 Plus, 100-240V, various plugs	MaxQ2000C02 Plus Promotion Package for USA, includes C02 resistant shaker (88881101), Universal Platform and Flask Clamp Starter Kit	MaxQ2000C02 Plus Promotion Package for EMEA & APAC, includes C02 resistant shaker (88881102), Universal Platform & Flask Clamp Starter Kit	
EU Plug	N/A	1	N/A	1	
UK Plug	N/A	1	N/A	1	
Universal Platform (AL)	N/A	N/A	1	1	
100ml Flask Clamp	N/A	N/A	2	2	
250ml Flask Clamps	N/A	N/A	4	4	
500ml Flask Clamps	N/A	N/A	4	4	
1L Flask Clamps	N/A	N/A	2	2	
2L Flask Clamps	N/A	N/A	2	2	
Data Port	1	1	1	1	

Universal (Aluminum) Platform Installation

1. Place the instrument on a level and firm surface; it must be capable of supporting the weight and load of the instrument, and the impact force generated by the motion. Prepare related parts (see Figure 1-1).



- 1) M4x14 cross recessed countersink head screws
- 2) Universal (aluminum) platform
- 3) Main body of CO₂ resistant shaker

Figure 1-1. Universal Platform Installation Diagram


2. Install the horizontal universal platform on the fixed plate of the Main body, ensure all four mounting holes in the universal platform and the four threaded mounting holes are in alignment.
3. Fasten the universal platform and the fixed plate with four M4x14 cross recessed countersink head screws.

Section 1

Unpacking and Installation

18 inch (Aluminum) Platform Installation

1. Place the instrument on a level and firm surface; it must be capable of supporting the weight and load of the instrument, and the impact force generated by the motion. Prepare related parts (see Figure 1-2).




- 1) M4x14 cross recessed countersink head screws
- 2) 18 inch (aluminum) platform
- 3) Main body of CO₂ resistant shaker

Figure 1-2. 18 inch Platform Installation Diagram

2. Install the horizontal universal platform on the fixed plate of the Main body, ensure all four mounting holes in the universal platform and the four threaded mounting holes are in alignment.
3. Fasten the universal platform and the fixed plate with four M4x14 cross recessed countersink head screws.

Rubber Mat Platform Installation

1. Place the instrument on a level and firm surface; it must be capable of supporting the weight and load of the instrument, and the impact force generated by the motion. Prepare related parts (see Figure 1-3).



- 1) Rubber mat
- 2) M4x10 cross recessed countersunk head screws
- 3) Platform
- 4) Main body of CO₂ resistant shaker

Figure 1-2. Universal Platform Installation Diagram


2. Install the horizontal platform on the fixed plate of the Main body; ensure all four mounting holes on the platform and the four threaded mounting holes are in alignment.
3. Fasten the platform and the fixed plate with four M4x10 cross recessed countersink head screws.
4. Place the rubber mat on the platform.

Section 1

Unpacking and Installation

Connections of the Instrument

Refer to Figure 1-4 when connecting the instrument.



- 1) Main body of CO₂ resistant shaker
- 2) Control box
- 3) Power adapter
- 4) Power cord
- 5) 4-20mA data port

Figure 1-4. Circuit Diagram

Note:

- 1) When disconnecting the control box from the main body, please cut off the power first. ▲
- 2) No special cable needed to connect with 4-20mA data port. ▲

Section 2 Overview

Specifications

Rotation Speed

Speed Range30 to 300 rpm
Speed Accuracy ± 1 rpm
Orbit Diameter $\varnothing 19$ mm
Motion TypeOrbital
Speed DisplayLED

Load

Maximum Load6kg @ ≤ 300 rpm

Timer

Timer Range0~99h 59min.
Hour DisplayLED
Minute DisplayLED

Dimensions (WxDxH)

Main body360 x 300 x 96mm (14.2 x 11.8 x 3.8 inch)
Control Box213 x 95 x 26mm (8.4 x 3.7 x 1.0 inch)

Weight

Net Weight (Main body)17.5 kg (38.58 lbs)
Net Weight (Control Box)0.3 kg (0.66 lbs)
Gross weight19 kg (41.9 lbs)

Power Supply

RequirementAC100 - 240V, 50/60Hz

Other

CertificationROHS, WEEE, cCSAus, CE Mark
Noise level<60dB
IP ClassIP21

Environmental Conditions

Application Environment Conditions: Indoor use

Temperature	4°C to 60°C
Altitude	≤2,000 m
Humidity20% to 99%
CO ₂ Concentration0% to 20%

Storage Environmental Conditions

Temperature5°C to 60°C
Altitude	≤2,000m
Humidity20% to 99%, Non-Condensing


Safety Tips

1. Unplug and unload before moving the unit.
2. This instrument uses a single-phase AC100 - 240V 50/60Hz power supply.
3. Remove the power supply when the instrument is in maintenance.
4. DO NOT clean the unit with flammable or corrosive liquid.
5. Use specified power supply.
6. Maximum load 6kg (13.23 lbs.) @≤300 rpm
7. To ensure proper operation, the unit must be free of surrounding obstructions.

Storing the Signal Cable

To store the signal cable, simply open the side plate on the main body. Refer to Figure 2-1. The signal cable can be placed inside.


Caution When storing the cable, pay attention to closing the side plate door, to avoid pinching your fingers. ▲



- 1) Main body of CO₂ resistant shaker
- 2) Slide plate
- 3) Control box

Figure 2-1. Storing the Signal Cable Diagram

Orbital Speed vs. Output Current Monitoring System




- 1) Control box
- 2) Data port

Figure 2-2. Output Current Monitoring System

Orbital Speed vs.
Output Current
Monitoring System
(Continued)

Table 2-1. Output current under different speed conditions

Speed (rpm)	Output current (mA)
0	4
30	4.8
300	20



Maximum Shaking Speed

Table 2-2. Maximum Shaking Speed on Rubber Platform in Relation to Flask Size

Flask Size		Max Speed (rpm)
Size	Liquid Volume	
50ml	≤50ml	200
100ml	≤100ml	150
150ml	≤150ml	150
200ml	≤200ml	150
250ml	≤250ml	150
500ml	≤500ml	150
1000ml	≤1000ml	150
2000ml	≤2000ml	150
3000ml	≤3000ml	150
5000ml	≤5000ml	150

Table 2-3. Maximum Shaking Speed and Flask Capacity on Universal Platform

Flask Size		Max Flask Capacity	Max Speed (rpm)
Size	Liquid Volume		
50ml	≤50ml	30	300
100ml	≤100ml	15	300
150ml	≤150ml	15	300
200ml	≤200ml	15	300
250ml	≤250ml	15	300
500ml	≤500ml	9	300
1000ml	≤500ml	4	260
	≤1000ml	4	240
2000ml	≤1000ml	4	240
	≤2000ml	4	220
3000ml	≤1500ml	3	250
	≤3000ml	3	220
5000ml	≤2500ml	1	220
	≤5000ml	1	200

Maximum Shaking Speed (Continued)

Note:

- 1) The flask clamps ($\leq 2L$) can be placed anywhere on the universal (aluminum) platform, even if the load is not balanced. ▲
- 2) Please refer to the “Liquid Volume” table in the specifications section for the maximum load. ▲

Table 2-4. Maximum Shaking Speed and Flask Capacity on 18 inch (aluminum) Platform

Flask Size		Max Flask Capacity	Max Speed (rpm)
Size	Liquid Volume		
50ml	$\leq 50\text{ml}$	49	210
100ml	$\leq 100\text{ml}$	25	210
150ml	$\leq 150\text{ml}$	25	210
200ml	$\leq 200\text{ml}$	18	210
250ml	$\leq 250\text{ml}$	18	210
500ml	$\leq 500\text{ml}$	10	190
1000ml	$\leq 500\text{ml}$	7	200
	$\leq 1000\text{ml}$	7	190
2000ml	$\leq 1000\text{ml}$	4	180
	$\leq 2000\text{ml}$	4	180
3000ml	$\leq 1500\text{ml}$	4	160
	$\leq 3000\text{ml}$	2	150
5000ml	$\leq 2500\text{ml}$	2	170
	$\leq 5000\text{ml}$	1	150

Note:


- 1) The flask clamps ($< 2L$) can be placed anywhere on the 18 inch (aluminum) platform, even if the load is not balanced. ▲
- 2) Please refer to the “Liquid Volume” table in the specifications section for the maximum load. ▲

Section 3 Operation


The unit has been designed with a “soft start/slow down” providing a smooth acceleration/deceleration during operation. This feature will decrease splashing of solutions from vessels during operation.

Warning The microprocessor control system may take 1 minute to synchronize the actual RPM of the universal (aluminum) platform to the desired RPM setting. Make sure that all the flasks and test tubes are firmly fixed on the universal platform and the flask clamps are fastened securely with the screws. ▲

Operations on the Control Panel



Operations on the Control Panel (Continued)

- 1) Display LED: Shows time or speed
- 2) RPM indicator light: LED shows current speed
- 3) TIME indicator light: LED shows current time.
- 4) WARN indicator light: Over-speed alarm.
- 5) RUN indicator light: The instrument is running.
- 6) RPM/TIME Key: While instrument is in operation, press to change time and speed display; when setting parameter, press to change minutes, hour or speed.
- 7) Up Arrow Key: Increases value. Hold for 3 seconds to increase rapidly.
- 8) Down Arrow Key: Decreases value. Hold for 3 seconds to decrease rapidly.
- 9)  Key: The power is On/Off.
- 10) Run/Stop Key: The instrument is running or stopped.

Speed and Time Settings

Speed setting

1. Press RPM/TIME key. The speed indicator light is lit.
2. Press up and down arrow keys to access the speed setting state, speed value starts flashing.
3. Press up and down arrow keys to change the speed value. Speed increases or decreases. The minimum unit is 1rpm.
4. During settings, the unit will reset to current values if no value is input after three seconds.
5. Press “Run/Stop” key to run or stop.

Time setting

1. Press RPM/TIME key. The time indicator light is lit.
2. Press up and down arrow keys to access the speed setting state. Speed value starts flashing.
3. Press up and down arrow keys to change the time value. Time increases or decreases. The minimum unit is 1 minute or 1 hour.
4. During settings, the unit will reset to current values if no value is input after three seconds.
5. Press “Run/Stop” key to run or stop.

Speed and Time Settings (Continued)

Notes

1. When the shaker is running, the LED display will switch between the speed and timer function every 30 seconds automatically. Press RPM/TIME key to manually toggle between the speed and timer value.
2. In “timer mode” the LED display will show the remaining running time.
3. In “continuous mode” the LED display will show elapsed run time.
4. After the run is complete, the LED display will read “end”. Press any key to place the shaker in standby mode.

Alarm System

Alarm System

1. During operation, the alarm will sound after 10 seconds if the unit has not begun shaking. In addition, the alarm will sound and stop for the speed over 350 rpm after 2 seconds.
2. Once the shaker begins to alarm, the RUN indicator light will not be visible and the WARN indicator light will flash, and the unit will stop running.

Power failure recovery

When the power supply is cut off suddenly to the shaker, resulting in a power failure, the unit will automatically run at the previous set parameter upon power restoration.

Section 3

Operation


Section 4 Application

Application on Lab Benches

To operate the shaker outside of a CO₂ incubator, refer to Specifications in Section 2.

Application in CO₂ Incubator

Place the main body of the shaker in the CO₂ incubator, and use the signal cable to connect the control box through the access port of the CO₂ incubator. Refer to Figure 4-1.



- 1) CO₂ Incubator
- 2) Control box
- 3) Power Cord
- 4) Main body of CO₂ Resistant Shaker

Figure 4-1. Application in CO₂ incubator

Warning

1. Never put the control box in the CO₂ incubator.
2. The back of control box is magnetized, and can be attached on the surface of the CO₂ incubator.
3. When the main body of CO₂ Resistant Shaker is in operation in the CO₂ incubator, keep a safe distance from the inner wall of the CO₂ incubator to avoid collisions.
4. The main body of the shaker should be connected to the control box wire through the access port of the CO₂ incubator.
5. If the cells are sensitive to vibrations, do not put them in the same incubator or on the shaker.

Section 5 **Safety Tips and Maintenance**

Safety Tips

1. Confirm local power supply voltage before use the unit.
2. The unit must not be plugged into a power supply extension with other pieces of equipment.
3. The data cable between the main body and control box cannot be extended.
4. Never modify the supplied power cables.
5. Service should only be performed by a qualified professional
6. In the following situations, the power supply must be unplugged:
 - When the unit is moved.
 - When the control box is opened.
 - When a fuse is replaced.
 - When the equipment is not functioning properly.
 - When the equipment is not in use.

Maintenance

Surfaces can be cleaned with mild detergent and water. Liquid should not be allowed to enter inside the unit.

Section 5

Safety Tips and Maintenance

Clean spill






If accidental spillage of liquids or other materials occur on the platform, shut down the machine and clean up the liquids or other materials immediately.

Note If liquid has entered the unit, remove the platform and clean it. ▲










1. Turn off the machine and unplug the power cord.
2. Remove the platform (with 4 screws).
3. Remove top plate (with 7 screws).
4. Use 70% ETOH to clean the inner area.
5. All allow the unit to dry for one hour, reassemble unit, and operate normally.

Section 6 Optional Accessories





Tabel 6-1. Universal (Aluminum) Platform

Description	Cat. No.	Dimensions (WxDxH) mm	Max. Qty.	Figure
50ml clamp	88881105	60×60×40	30	
100ml clamp	88881106	60×65×45	15	
150ml clamp	88881107	80×80×50	15	
200ml clamp	88881108	80×80×55	15	
250ml clamp	88881109	90×90×60	15	
500ml clamp	88881110	115×115×70	9	
1000ml clamp	88881111	130×130×75	4	
2000ml clamp	88881112	170×170×100	4	
3000ml clamp	88881113	200×200×140	2	
5000ml clamp	88881114	245×245×140	1	
500ml Infusion bottle clamp	88881115	100×100×95	6	
1000ml Infusion bottle clamp	88881116	125×125×125	4	
Funnel Clamp	88881117	N/A	2	
Fixed Tube rack 40×Ø14mm	88881118	265x120x140	2	
Fixed Tube rack 40×Ø16mm	88881119	265x120x140	2	






Tabel 6-1. Universal (Aluminum) Platform (Continued)

Description	Cat. No.	Dimensions (WxDxH) mm	Max. Qty	Figure
Adjustable Tube rack 40×Ø14mm	88881120	285x125x175	2	
Adjustable Tube rack 40×Ø16mm	88881121	285x125x175	2	
Universal (Aluminum) platform	88881122	360x300x5	1	
Rubber Mat platform	88881123	360x305x10	N/A	
96-well microplate holder	88881124	130x90x15	6	
Spring Wire Rack	88881125	350X240X80	1	
Sticky Mat	88881126	140x140	4	
Control Box	88881127B	213x95x26	1	
General Power Adaptor	88881128	N/A	1	
US Plug	88881129	N/A	1	










Tabel 6-1. Universal (Aluminum) Platform

Description	Cat. No.	Dimensions (WxDxH) mm	Max. Qty	Figure
CN Plug	88881130	N/A	1	
EU Plug	88881131	N/A	1	
UK Plug	88881132	N/A	1	
Flask Clamp Starter Kit	88881133	N/A	1	100ml flask clamp x 2 250ml flask clamps x 4 500ml flask clamps x 4 1L flask clamps x 2 2L flask clamp x 2
Data port	88881181	41X41X41	1	





Tabel 6-2. 18 inch (Aluminum) Platform

Description	Cat. No.	Dimensions (WxDxH) mm	Max. Qty.	Figure
50ml clamp	88881105	60×60×40	49	
100ml clamp	88881106	60×65×45	25	
150ml clamp	88881107	80×80×50	25	
200ml clamp	88881108	80×80×55	18	
250ml clamp	88881109	90×90×60	18	
500ml clamp	88881110	115×115×70	10	
1000ml clamp	88881111	130×130×75	7	
2000ml clamp	88881112	170×170×100	4	
3000ml clamp	88881113	200×200×140	4	
5000ml clamp	88881114	245×245×140	2	
500ml Infusion bottle clamp	88881115	100×100×95	12	
1000ml Infusion bottle clamp	88881116	125×125×125	9	
Funnel Clamp	88881117	N/A	6	
Fixed Tube rack 40×Ø14mm	88881118	265x120x140	4	
Fixed Tube rack 40×Ø16mm	88881119	265x120x140	4	

Tabel 6-2. 18 inch (Aluminum) Platform (Continued)

Description	Cat. No.	Dimensions (WxDxH) mm	Max. Qty	Figure
Adjustable Tube rack 40xØ14mm	88881120	285x125x175	4	
Adjustable Tube rack 40xØ16mm	88881121	285x125x175	4	
18 inch (Aluminum) platform	88881122B	457x457x6	1	
Rubber Mat platform	88881123	360x305x10	N/A	
96-well microplate holder	88881124	130x90x15	10	
Spring Wire Rack	88881125	350X240X80	1	
Sticky Mat	88881126	140x140	9	
Control Box	88881127B	213x95x26	1	
General Power Adaptor	88881128	N/A	1	
US Plug	88881129	N/A	1	

Tabel 6-2. 18 inch (Aluminum) Platform (Continued)

Description	Cat. No.	Dimensions (WxDxH) mm	Max. Qty	Figure
CN Plug	88881130	N/A	1	
EU Plug	88881131	N/A	1	
UK Plug	88881132	N/A	1	
Flask Clamp Starter Kit	88881133	N/A	1	100ml flask clamp x 2 250ml flask clamps x 4 500ml flask clamps x 4 1L flask clamps x 2 2L flask clamp x 2
Data port	88881181	41X41X41	1	

Section 7 Troubleshooting

To troubleshoot the unit, refer to the table below. If your problem persists, contact your local sales/service representative.

Table 7-1. Troubleshooting

Error	Cause	Solution
Indicator off No reaction	Power disconnected	Connect the power
	Power adapter failure	Replace power adapter
	Switched off	Switch on
No rotating of platform	Sensor failure	Contact local sales/service
	Main-board failure	Contact local sales/service
Loud noise	The unit has been placed on uneven surface	Place the unit on level and sturdy surface.
	Platform loose	Fasten platform
	Accessories loose	Fasten accessories
Other	Keep record for maintenance	

Section 7
Troubleshooting

THERMO FISHER SCIENTIFIC STANDARD PRODUCT WARRANTY

The Warranty Period starts two weeks from the date your equipment is shipped from our facility. This allows for shipping time so the warranty will go into effect at approximately the same time your equipment is delivered. The warranty protection extends to any subsequent owner during the first year warranty period.

During the first two (2) years, component parts proven to be non-conforming in materials or workmanship will be repaired or replaced at Thermo's expense, labor included. Installation and calibration are not covered by this warranty agreement. The Technical Services Department must be contacted for warranty determination and direction prior to performance of any repairs. Expendable items, glass, filters and gaskets are excluded from this warranty.

Replacement or repair of components parts or equipment under this warranty shall not extend the warranty to either the equipment or to the component part beyond the original warranty period. The Technical Services Department must give prior approval for return of any components or equipment. At Thermo's option, all non-conforming parts must be returned to Thermo Fisher Scientific postage paid and replacement parts are shipped FOB destination.

THIS WARRANTY IS EXCLUSIVE AND IN LIEU OF ALL OTHER WARRANTIES, WHETHER WRITTEN, ORAL OR IMPLIED. NO WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE SHALL APPLY. Thermo shall not be liable for any indirect or consequential damages including, without limitation, damages relating to lost profits or loss of products.

Your local Thermo Sales Office is ready to help with comprehensive site preparation information before your equipment arrives. Printed instruction manuals carefully detail equipment installation, operation and preventive maintenance.

If equipment service is required, please call your Technical Services Department at 1-800-438-4851 (USA and Canada) or 1-740-373-4763. We're ready to answer your questions on equipment warranty, operation, maintenance, service and special application. Outside the USA, contact your local distributor for warranty information.

Rev. 0 9/13

Section 8

Warranty Information

thermoscientific.com

© 2021 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific and its subsidiaries. Specifications, terms and pricing are subject to change. Not all products are available in all countries. Please consult your local sales representative for details.

Thermo Fisher Scientific (Asheville) LLC

401 Millcreek Road
Marietta, Ohio 45750
United States

Thermo
S C I E N T I F I C
A Thermo Fisher Scientific Brand